

CERTIFICATE OF ANALYSIS

PRODUCT	PB5220E	COL.-C/NA-ÄSCULIN-NB/BRILLIANCE (TM) UTI
LOT NUMBER	3489674	
EXPIRY DATE	2022.06.21	
TEST DATE	2022.05.11	
REPORTING DATE	2022.05.16	

General Characteristics Side 1	Results	Specification
Colour	Conforms	Flame red
Appearance	Conforms	Opaque
pH	7.2	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check

Microbiological Performance Side 1	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®25923	<10000	Conforms	Good growth, white shiny colonies
Streptococcus agalactiae ATCC®13813	<10000	Conforms	Good growth, grey shiny cols, Aesculin -
Enterococcus faecalis ATCC®29212	<10000	Conforms	Good growth, light grey cols, Aesculin +
Escherichia coli ATCC®25922	>100,000	Conforms	No growth

General Characteristics Side 2	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Opaque
pH	6.9	6.6 -7.0
Packaging / Presentation	Conforms	Label & Print check

Microbiological Performance Side 2	Control c.f.u	Test Result	Specification
Escherichia coli ATCC®25922	<10000	Conforms	Good growth, rose shiny colonies, Indol+
Proteus mirabilis ATCC®29906	<10000	Conforms	Good growth, cream colonies, Indol -
Enterococcus faecalis ATCC®29212	<10000	Conforms	Good growth, turquoise shiny colonies
Klebsiella oxytoca ATCC®13182	<10000	Conforms	Good growth, large blue shiny colonies
Staphylococcus aureus ATCC®25923	<10000	Conforms	Good growth, white colonies

3 Day - Contamination Check	Results	Specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

The quality control methods meet requirements of ISO 11133.

CERTIFICATE OF ANALYSIS

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkks according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.